



Innovative Bioanalysis
3188 Airway Ave Suite - D
Costa Mesa CA, 92626
www.InnovativeBioanalysis.com
Email: Albert.Brockman@innovativebioanalysis.com

Methicillin Resistant S. Aureus Reduction Through Ionization

CLIENT: ACA/IAE

PROJECT: ACA/IAE-ION-MRSA-01

REPORT NUMBER: ACA/IAE-ION-MRSA-01

PRODUCT: ACA-RN-0001 and ACA4800GU-1, Powered by GPS NPBI™ Technology

CAP LIC NO: 886029801

CLIA LIC NO: O5D0955926

STATE ID: CLF 00324630

REPORT DATE: 01/04/2021

CHALLENGE ORGANISIM(S):

- **Methicillin Resistant Staphylococcus Aureus**



EXPERIMENT SUMMARY

ACA/IAE supplied both a **ACA-RN-0001** and **ACA4800GU-1** system for testing purposes to determine efficacy against Methicillin Resistant Staphylococcus Aureus. When operational the units are intended to neutralize bacteria on surfaces and in the air to sanitize enclosed areas.

The ACA-RN-0001 and ACA4800GU-1 are designed to deactivate viral and bacterial pathogens on surfaces and in the air to sanitize enclosed areas. This study was to evaluate the efficacy of one bacterial strain referred to as MRSA in a large setting. Both the ACA-RN-0001 and ACA4800GU-1 have been tested and evaluated to be consistent in production of equal ions in the test chamber. Both systems have been used in the following test procedures and the results of each system are identical in performance and ion production.

BACTERIA

BEI Resources Catalog number NR-41886 (CoA attached) Methicillin Resistant S. Aureus was cultured by plating the thawed broth on Tryptic Soy Agar and allowed to incubate at 32°C with 5% CO₂ for 24 hours. A single isolate colony was harvested and introduced to a tryptic soy broth and allowed to incubate at 32°C for an additional 24 hours. Upon completion of the incubation period, bacteria were harvested and rinsed 3 times in phosphate buffered saline. A 1 to 10 dilution was made by removing 1 mL of inoculated tryptic soy broth and adding it 9 mL of phosphate buffered saline. This solution was further diluted to a final concentration of 1:100.

CONTROL SUMMARY:

For the control section one AIC2 Air Ion counter was placed in the center of the of the testing chamber for 5 minutes prior to the control test. The natural state of ions was counted, and little fluctuations were observed. Ion counts were recorded every 0.5 seconds and the average for the duration of the test was 73 ions per cm³ without the ionization unit running.

5 sterile dishes containing organisms were provided by the lab staff, labeled with time point designation and organism. Dishes were placed on a table inside the room and the door closed to prevent outside environmental contaminants. Swabs were taken at the pre-defined time points of 0 Minutes, 10 minutes, 20 minutes, 30 minutes and 60 minutes for MRSA and all swabs were sealed after collection and provided to lab staff. The door to the chambers remained closed the entirety of the test and all air entering the test chamber was filtered through a HEPA filter.



MATERIALS AND EQUIPMENT:

- **Certified Biological Safety Cabinet**
- **Micropipette and sterile disposable aerosol resistant tips – 20uL, 200 uL, 1000uL**
- **Microscope**
- **Tubes for dilution**
- **Hemocytometer with cover slip**
- **Tryptic Soy Broth**
- **Tryptic Soy Agar**
- **10 uL Inoculation Loops**
- **CO2 Incubator set at 34°C**

EXPERIMENTAL SUMMARY: T0, T10, T20, T30, T60


The test stage can be seen in the design diagram and consisted of a metal and laminate safety test chamber measuring, 20'W x 8'H x 8'D with sealed seams was used for the testing site. An equalizing vent was on opened on the side of the test chamber to allow natural air to enter and exit from the chamber. The air temperature fluctuated slightly through the test and ranged from 72.41F to 73.77F. During the control testing and the viral load tests the temperature fluctuation was consistent. The ambient humidity inside the test chamber averaged 53.7%.

- Behind the rowed test site there was an AIC2 Air Ion Counter continually logging the negative ion count.
- Test row contained 5 round petri dishes provided by the lab inoculated with 1mL of MRSA with a concentration of 33,152 CFU/mL.
- All sample dishes were labeled with their bacterium and the time point they were to be used with. 1 sample swab was taken from each dish, as well as a swab collected for residual bacteria at 0-minute timepoint, 10-minute time point, 20-minute time point, and 30-minute time point and 60-minute time point.
- The ion system was attached to a variable speed fan and angled up at a 45-degree angle to allow the ions to cascade down throughout the room
- Fan speeds were adjusted until the desired concentration was reached prior to exposing test samples. All samples were brought into the testing chamber sealed and removed from the testing chamber sealed.
- Upon testing completion, samples were provided to lab staff for further review.


INNOVATIVE BI•ANALYSIS

creating solutions | getting results

DESIGN LAYOUT:



TEST AVERAGE 27K PER CUBIC CENTIMETER



INNOVATIVE BI·ANALYSIS

creating solutions | getting results

RESULTS:

<u>CONTROL RESULTS</u>			<u>EXPERIMENT RESULTS</u>		
<u>Timepoint</u>	<u>CFU/mL</u>	<u>% Reduction</u>	<u>Timepoint</u>	<u>CFU/mL</u>	<u>% Reduction</u>
0 Min	32,212 CFU/mL	1.2	0 Min	31,816 CFU/mL	2.4
10 Min	30,226 CFU/mL	6.1	10 Min	24,358 CFU/mL	23.4
20 Min	29,484 CFU/mL	8.4	20 Min	13,166 CFU/mL	58.6
30 Min	27,842 CFU/mL	13.4	30 Min	5,534 CFU/mL	82.6
60 Min	24,618 CFU/mL	23.6	60 Min	826 CFU/mL	97.4

CONCLUSIONS:

The overall reduction of controls (natural degradation) versus experiment shows the increased degradation of MRSA in an ionized environment. It is important to note the natural degradation at each time point when considering the effectiveness of the ionized environment. In conclusion, utilization of the ACA-RN-0001 and ACA4800GU-1, in the test environment significantly decreased the concentration of MRSA in a 60-minute time by 97.4% overall and 73.8% better than natural degradation.

INNOVATIVE BIOANALYSIS

creating solutions | getting results

DocuSigned by:

Dr. Dana Yee M.D.

7D5A69A0907947B...

1/4/2021

Dr. Dana Yee M.D
Clinical Pathologist and Medical Director

Date

DocuSigned by:

Sam Kabbani

8B4B282DF4B34A3...

1/4/2021

Sam Kabbani, MS, BS, MT(ASCP), CLS
Chief Scientific Officer, Innovative Bioanalysis

Date

DocuSigned by:

Albert Brockman

06DF5C77A0D2400...

1/4/2021

Albert Brockman
Chief Biosafety Officer, Innovative Bioanalysis

Date

DocuSigned by:

Kevin Noble

5DF2797BAA78421...

1/4/2021

Kevin Noble
Chief Operating Officer, Innovative Bioanalysis

Date

INNOVATIVE BI•ANALYSIS

creating solutions | getting results

CERTIFICATE OF ANALYSIS

***Staphylococcus aureus*, Strain M0602 (MRSA)**

Catalog No. NR-41886

Product Description: *Staphylococcus aureus* (*S. aureus*), strain M0602 was isolated in 2004 from a blood sample of an adult male. *S. aureus*, strain M0602 is a methicillin-resistant *S. aureus* (MRSA) strain.

Lot¹: 62173712

Manufacturing Date: 14NOV2013

TEST	SPECIFICATIONS	RESULTS
Phenotypic Analysis Cellular morphology Colony morphology ² Motility (wet mount) Oxacillin resistance ³ Hemolysis ² Biochemical characterization: Analytical profile index (API Staph [®]) ⁴ Catalase	Gram-positive cocci Report results Report results Resistant Report results Consistent with <i>S. aureus</i> Positive	Gram-positive cocci Circular, low convex, entire, smooth and gray (Figure 1) Non-motile Resistant β-hemolytic Consistent with <i>S. aureus</i> Positive
Genotypic Analysis Sequencing of 16S ribosomal RNA gene (~ 910 base pairs) Riboprinter [®] Microbial Characterization System	Consistent with <i>S. aureus</i> Consistent with <i>S. aureus</i>	Consistent with <i>S. aureus</i> Consistent with <i>S. aureus</i>
Viability (post-freeze)²	Growth	Growth

¹NR-41886 was produced by inoculation of the deposited material into Tryptic Soy broth and grown 24 hours at 37°C in an aerobic atmosphere. Broth inoculum was added to Tryptic Soy agar with 5% defibrinated sheep blood kolles which were grown 24 hours at 37°C in an aerobic atmosphere to produce this lot. Purity of this lot was assessed for 7 days under propagation conditions.

²24 hours at 37°C and aerobic atmosphere on Tryptic Soy agar with 5% defibrinated sheep blood

³24 hours at 30°C and aerobic atmosphere on Mueller Hinton agar containing 6 µg/mL oxacillin and 4% NaCl (Oxacillin Screen Agar BBL™ 221952)

⁴The API Staph[®] identification includes a test for lysostaphin resistance that was not completed for NR-41886. NR-41886 is identified as *S. aureus* by the API Staph[®] test regardless of the lysostaphin resistance test results.

Figure 1



Date: 15 FEB 2014

Signature:

Title: Technical Manager, BEI Authentication or designee

ATCC[®], on behalf of BEI Resources, hereby represents and warrants that the material provided under this certificate has been subjected to the tests and procedures specified and that the results described, along with any other data provided in this certificate, are true and accurate to the best of ATCC[®]'s knowledge.

ATCC[®] is a trademark of the American Type Culture Collection.

You are authorized to use this product for research use only. It is not intended for human use.

